





# Wide Range Measurement From 1 to 20 000 Hz







With the auto store function included as standard, as well as a timer function and external power supply support, the NL-62 is ideal for continuous measurement. Designed for intuitive ease of use, there is no more need to consult the manual during a measurement. The large 3-inch color screen is bright and easy to read. Sudden rainfall is also no problem, thanks to the water-resistant construction. Using the optional octave and 1/3 octave band real-time analysis program NX-62RT (under development), the unit can even operate as a frequency analyzer. The High-Precision Sound Level Meter NL-62 supports all your measurement needs.



255 mm

10 inch

Large color LCD screen

Equipped with non-slip rubber grips

Three-inch LCD screen with a touch panel High resolution screen is easy to see indoors or outdoors and even in the dark.



(φ2.5 mono jack)

(Full scale)


START/STOP

LIGHT


PAUSE/CONT

## No paper manual is needed.


User instructions and a help function can be easily accessed on the device.



Measurement Display (Level-Time graph)



Measurement Display (low-frequency sound)



Parameter Screen



Menu screen



Help screen

#### Water-resistant (Except for the microphone)

Guaranteed water-resistant to at least level IP54 (resistant to spraying water). Helps reduce failures caused by sudden rain showers.



## Use of rechargeable batteries

In these new models it is possible to use rechargeable batteries which make these meters environmentally-friendly. 16 hour continuous measurement is possible (when using eneloop® or dry alkaline batteries).



- Please use the dedicated charger to charged eneloop® batteries
- When using eneloop batteries, please read the eneloop® battery instruction manual.
   eneloop® is a registered trademark of Panasonic group.

## Continuous detailed measurements for one month

This meter can be used to conduct long-term measurements, such as environmental measurements. (If an AC adapter is used)

Duration of recording

1000 h (approx. one month)


Previous model 200 h (approx. one week)

Example of detailed recording


If the  $L_p$  is measured at 100 ms intervals and the  $L_{eq}$  is simultaneously measured at 10 min intervals over a 24 h period, the total size of accumulated data is approximately 74 MB (reference value)

Functionality can be extended by a range of options


Add long-term data recording capability and frequency analysis function



1/3 octave band analysis screen (low range)



FFT analysis screen (x40)



Data management screen of AS-60 software

### **Program function list**

#### Auto store function


This function enables continuous measurement in  $L_p$  mode (instantaneous SPL) and  $L_{eq}$  mode (equivalent continuous SPL) to be conducted simultaneously.

Total measuring time of Auto store function

Up to 1000 h


Equipped with a timer function

Lp mode (instantaneous SPL) and Leq mode (equivalent continuous SPL) concept



#### Comparator function

This function turns on when the open collector output exceeds the set value (max. applied voltage 24 V, max. current 60 mA, allowable dissipation 300 mW).



#### Continuous data output function

This function enables the continuous acquisition of instantaneous values and processed values during both USB and RS-232C communication.

This is a convenient function for users who can design their own control programs, where data has to be transferred continuously from the sound level meter to the computer.

## **Optional program function list**

## Octave, 1/3 octave real-time analysis program NX-62RT



The NX-62RT is supplied on the 512 MB SD card. The 512 MB SD card can be used as a memory card after installing the program.




By adding the NX-62RT program to the NL-62, octave band and 1/3 octave band real-time analysis can be realized. Saved analysis results can be loaded and shown in an overlay graph display together with current analysis data. NC curve graph display and NC value calculation/display are also possible.




1/3 octave band analysis screen (low range)



1/3 octave band analysis screen (high range)



Overlay analysis screen



1/3 octave band analysis screen (combined bands)



Measurement screen (Level-Time graph)

## Waveform recording program NX-42WR



The NX-42WR is supplied on the 2 GB SD card. The 2 GB SD card can be used as a memory card after installing the program.



This function enables users to record sounds and processing sound to levels simultaneously. Recorded data can be played on computer and used for frequency analysis.

(Uncompressed waveform WAVE file)

Sampling at 48 kHz, 24 kHz, 12 kHz, Selection of 24 bit or 16 bit

Maximum recording time (16 bit)


Memory card Sampling frequency	512 MB	2 GB
48 kHz	1 h	4 h
24 kHz	2 h	8 h
12 kHz	4 h	16 h

## Reverberation Time Measurement Program NX-42RV




The NX-42RV is supplied on the 512 MB SD card. The 512 MB SD card can be used as a memory card after installing the program.


By adding the NX-42RV program to the NL-52/42, reverberation time measurements can be performed. The measurement method is the interrupted noise method. This program allows storage of reverberation time decay curves, T20/T30 calculation, Txx calculation (reverberation time calculation based on a user-defined interval) and averaged reverberation time results displayed on the SLM screen.




Measuring screen (graph)



Measuring screen (numeric)



Reverberation time decay curve screen



Result screen (T20/T30/Txx)

# FFT analysis program NX-42FT




The NX-42FT is supplied on the 512 MB SD card. The 512 MB SD card can be used as a memory card after installing the program.




By adding the NX-42FT program to the NL-62, FFT analysis can be performed. The analysis frequency range is 20 kHz, with 8 000 spectrum lines (200 displayed). Saved analysis results can be loaded and shown in an overlay graph display together with current analysis data. Maximum zoom ratio is x40, and the top list screen can show up to 20 lines.




Analysis screen (x1)




Analysis screen (x40)



Overlay analysis screen




Linear average screen



Top list screen

#### **System construction**



#### Peripheral devices

# All-weather windscreen WS-15



This windscreen is designed for outdoor installations. It helps to reduce wind noise and is equipped with rainproof features that satisfy the IPX3 water-resistant specifications. It is used with a microphone extension cable.

(Mounting adapter WS15006 required separately)

# Rain-protection windscreen **WS-16**



This screen protects the microphone against rain for a short period of time.
The rainproof performance of this windscreen is designed to satisfy the IPX3 water-resistant specifications.

# Sound calibrator NC-74



This Sound calibrator conforms to IEC 60942 (JIS C 1515), Class 1, providing a level of performance sufficient for calibrating the precision sound level meter.

Specifications
Nominal acoustic pressure level 94 dB
Nominal frequency 1 kHz

#### Tripod

This stand can be used for general acoustic measurements. The sound level meter and microphone can be mounted on the stand.




(For All-weather windscreen WS-15, use of ST-81 is recommended.)

# Waveform analysis software AS-70

This software allows you to load stored WAVE files from a RION sound level meter, vibration meter or data recorder. Octave, 1/3 octave, and FFT analyses can then be performed. Playback of the real sound files is also possible.

Specifications		
Waveform analysis	Calculations	Maximum value, Minimum value, Average value, RMS, Variance,
		Differential and integral calculus, HPF, LPF
Frequency weighting	ng	Z, A, C, G, C to A, L <sub>v</sub> (vertical) (JIS C 1510), L <sub>v</sub> (horizontal) (JIS C 1510)
FFT analysis	Analysis points	32 to 65 536 points
	Display data	Power spectrum, Power spectral density, Spectrogram
Time weighting		10 ms, F, 630 ms, S, 10 s
Octave band	Applicable standards	IEC 61260 Class 1 (JIS C 1514 Class 1)
analysis	Analysis frequency	Octave band 0.5 Hz to 16 kHz (16 bands)
	range	1/3 octave hand 0.4 Hz to 20 kHz (48 hands)



Frequency analysis screen

CPU Intel Core™2 Duo 2 GHz or higher

RAM 2 GB or more (4 GB recommended)

HDD 20 GB free or more (100 GB or more recommended)

DISPLAY XGA (1 024 × 768) or more OS Microsoft Windows XP Professional 32 bit, 7 Professional 32 bit / 64 bit, 8 Pro 32 bit/64 bit

Recommended computer specifications

## Complete software for environmental measurements


#### Data management software for environmental measurement AS-60

Data management software for environmental measurement AS-60 enables the graph display of measurement data, arithmetic processing, excluded sound processing, preparation of reports, output of files, and playback of real sound files.


- Reports easy to prepare
- Simultaneous display of multiple data items (up to 8 data items)
- Data stored in a data recorder can Data combination be loaded (CSV file for DA-40 Viewer)

trial version now available on

our website



#### Data management software for environmental measurement AS-60RT (Includes the octave and 1/3 octave data management software)



#### Adds support for handling octave band analysis data to AS-60

AS-60RT is for managing NX-62RT/42RT or NA-28 data on a computer.



## Data management software for environmental measurement AS-60VM (Includes the vibration level data management software)

Adds support for handling data measured with VM-53A to AS-60

Supported models VM-53A\* \*Only auto store data are supported.

#### Specifications

			NL-62
Applic	cable	standards	IEC 61672-1: 2002 Class 1
			ISO 7196: 1995
			ANSI S1.4-1983 Type 1
			ANSI S1.4A-1985 Type 1
			ANSI S1.43-1997 Type 1
			JIS C 1509-1: 2005 Class 1
			CE Marking (EMC Directive 2004/108/EC, Low Voltage Directive 2006/95/EC),
			WEEE Directives, Chinese RoHS (export model for China only)
Measi	urem	ent functions	Simultaneous measurement of the following items, with selected time
	u. 0		weighting and frequency weighting
Pro	ncess	sing (main ch)	Instantaneous sound pressure level: Lp
10	0000	sing (main cin)	Equivalent continuous sound pressure level: Lea
			·
			Sound exposure level: $L_E$
			Maximum sound pressure level: L <sub>max</sub>
			Minimum sound pressure level: L <sub>min</sub>
<u> </u>			Percentile sound levels: L <sub>N</sub> (0.1 to 99.9 %, 0.1-increment steps, max. 5 values)
Pro	ocess	sing (sub ch)	Instantaneous sound pressure level: Lp
Add	ditior	nal processing	One of the following can be selected:
			C-weighted equivalent continuous sound level: L <sub>Ceq</sub>
			G-weighted average sound level: L <sub>Geq</sub>
			C-weighted peak sound level: L <sub>Cpeak</sub>
			Z-weighted peak sound level: L <sub>Zpeak</sub>
			Power average of max. level in time weighted sound level interval L <sub>Atm5</sub>
			I-time-weighted average sound level: LAIeq
			Max. value of I-time-weighted average sound level: LAImax
			*Because additional processing frequency characteristics are linked to sub channel
			frequency characteristics, L <sub>Alm5</sub> , L <sub>AIeq</sub> , L <sub>AImax</sub> can be selected when A characteristics are selected for sub channel. When C, G, or Z characteristics are
	. 1		selected, $L_{Ceq}$ and $L_{Cpeak}$ , $L_{Geq}$ , and $L_{Zpeak}$ can be selected for additional processing.
Microph	ŀ	,,	UC-59L
		Sensitivity level	-27 dB
Meası	urem	ent range	A-weighting: 25 dB to 138 dB
			C-weighting: 33 dB to 138 dB
			G-weighting: 43 dB to 138 dB
			Z-weighting: 50 dB to 138 dB
			C-weighting peak sound level: 60 dB to 141 dB
			Z-weighting peak sound level: 65 dB to 141 dB
Inhere	ent	A-weighting	17 dB or less
noise	ŀ	C-weighting	25 dB or less
110100	ı		35 dB or less
	ı	G-weighting	
-		Z-weighting	42 dB or less
		range	1 Hz to 20 kHz
	_	weighting	A, C, G and Z
Time v	weig	hting	F (Fast) and S (Slow), I (Impulse) and 10 s
Level		ιο.	
Ror	rang	<u> </u>	Single range (Linearity range: 113 dB)
Dal		n display range max	
_	r graph		
Swit	r graph itching	n display range max	Max. 110 dB (20 to 130 dB)
Swit RMS (	r graph itching dete	n display range max of bar graph display ction circuit	Max. 110 dB (20 to 130 dB) Set the upper/ lower limit in 10 dB increments. Digital processing method
Swit RMS (	r graph itching dete	n display range max of bar graph display ction circuit	Max. 110 dB (20 to 130 dB)  Set the upper/ lower limit in 10 dB increments.  Digital processing method  20.8 μs (Lρ, Leq, LE, Lmax, Lmin, Lpeak : sampling frequency: 48 kHz)
Swit RMS of Samp	r graph itching detections	n display range max of bar graph display ction circuit cycle	Max. 110 dB (20 to 130 dB)  Set the upper/ lower limit in 10 dB increments.  Digital processing method  20.8 μs (Lρ, Leq, LE, L <sub>max</sub> , L <sub>min</sub> , L <sub>peak</sub> : sampling frequency: 48 kHz)  100 ms (L <sub>N</sub> )
Swit RMS of Samp	r graph itching detections	n display range max of bar graph display ction circuit cycle	Max. 110 dB (20 to 130 dB)  Set the upper/ lower limit in 10 dB increments.  Digital processing method  20.8 μs (Lρ, Leq, LE, L <sub>max</sub> , L <sub>min</sub> , L <sub>peak</sub> : sampling frequency: 48 kHz)  100 ms (L <sub>N</sub> )  Electrical calibration performed according to IEC and JIS standards, using
Swit RMS of Samp Calibr	deter	n display range max of bar graph display ction circuit cycle	Max. 110 dB (20 to 130 dB)  Set the upper/ lower limit in 10 dB increments.  Digital processing method  20.8 μs (Lρ, Leq, LE, Lmax, Lmin, Lpeak : sampling frequency: 48 kHz)  100 ms (L <sub>N</sub> )  Electrical calibration performed according to IEC and JIS standards, using internally generated signals: acoustic calibration performed with the NC-74.
Swit RMS of Samp Calibr	deter	n display range max of bar graph display ction circuit cycle	Max. 110 dB (20 to 130 dB)  Set the upper/ lower limit in 10 dB increments.  Digital processing method  20.8 µs (L <sub>P</sub> , L <sub>eq</sub> , L <sub>E</sub> , L <sub>max</sub> , L <sub>min</sub> , L <sub>peak</sub> : sampling frequency: 48 kHz)  100 ms (L <sub>N</sub> )  Electrical calibration performed according to IEC and JIS standards, using internally generated signals: acoustic calibration performed with the NC-74.  Windscreen correction:
Swit RMS of Samp Calibr	deter	n display range max of bar graph display ction circuit cycle	Max. 110 dB (20 to 130 dB)  Set the upper/ lower limit in 10 dB increments.  Digital processing method  20.8 µs (Lp, Leq, Le, Lmax, Lmin, Lpeak : sampling frequency: 48 kHz)  100 ms (L <sub>N</sub> )  100 ms (L <sub>N</sub> )  Electrical calibration performed according to IEC and JIS standards, using internally generated signals: acoustic calibration performed with the NC-74.  Windscreen correction:  Compliant with IEC 61672-1 and JIS C 1509-1 standards when the
Swit RMS of Samp Calibr	deter	n display range max of bar graph display ction circuit cycle	Max. 110 dB (20 to 130 dB)  Set the upper/ lower limit in 10 dB increments.  Digital processing method  20.8 \( \mu \) \( \mu \), \( L_{\text{eq}}, L_{\text{E}}, L_{\text{max}}, L_{\text{min}}, L_{\text{peak}} \) : sampling frequency: 48 kHz)  100 ms \( L_{\text{N}} \)  100 ms \( (L_{\text{N}} \))  Electrical calibration performed according to IEC and JIS standards, using internally generated signals: acoustic calibration performed with the NC-74.  Windscreen correction:  Compliant with IEC 61672-1 and JIS C 1509-1 standards when the windscreen is installed.
Swit RMS of Samp Calibr	deter	n display range max of bar graph display ction circuit cycle	Max. 110 dB (20 to 130 dB)  Set the upper/ lower limit in 10 dB increments.  Digital processing method  20.8 \( \mu \) (L\( \nu \), L\( \n
Swit RMS of Samp Calibr	deter	n display range max of bar graph display ction circuit cycle	Max. 110 dB (20 to 130 dB)  Set the upper/ lower limit in 10 dB increments.  Digital processing method  20.8 \( \mu \) (L\( \nu \), L\( \n
Swit RMS of Samp Calibr	deter	n display range max of bar graph display ction circuit cycle	Max. 110 dB (20 to 130 dB)  Set the upper/ lower limit in 10 dB increments.  Digital processing method  20.8 \( \mu \) (L\( \nu \), L\( \n
Switt RMS ( Samp Calibr	r graph deterior oling of ration	n display range max of bar graph display cition circuit cycle	Max. 110 dB (20 to 130 dB)  Set the upper/ lower limit in 10 dB increments.  Digital processing method  20.8 μs (Lp, Leq, LE, Lmax, Lmin, Lpeak: sampling frequency: 48 kHz)  100 ms (L <sub>N</sub> )  Electrical calibration performed according to IEC and JIS standards, using internally generated signals: acoustic calibration performed with the NC-74.  Windscreen correction:  Compliant with IEC 61672-1 and JIS C 1509-1 standards when the windscreen is installed.  Diffuse sound field correction:  Correction of frequency characteristics in order to comply with standards
Switt RMS ( Samp Calibr	r graph deterior oling of ration	n display range max of bar graph display cition circuit cycle	Max. 110 dB (20 to 130 dB)  Set the upper/ lower limit in 10 dB increments.  Digital processing method  20.8 μs (L <sub>P</sub> , Leq, LE, L <sub>max</sub> , L <sub>min</sub> , L <sub>peak</sub> : sampling frequency: 48 kHz)  100 ms (L <sub>N</sub> )  Electrical calibration performed according to IEC and JIS standards, using internally generated signals: acoustic calibration performed with the NC-74.  Windscreen correction:  Compliant with IEC 61672-1 and JIS C 1509-1 standards when the windscreen is installed.  Diffuse sound field correction:  Correction of frequency characteristics in order to comply with standards (ANSI S1.4) in diffuse sound field.
Switt RMS of Samp Calibr Correct	r graph detering detering ration	n display range max of bar graph display cition circuit cycle	Max. 110 dB (20 to 130 dB)  Set the upper/ lower limit in 10 dB increments.  Digital processing method  20.8 µs (Lp, Leq, Le, Lmax, Lmin, Lpeak: sampling frequency: 48 kHz)  100 ms (Ln)  Electrical calibration performed according to IEC and JIS standards, using internally generated signals: acoustic calibration performed with the NC-74.  Windscreen correction:  Compliant with IEC 61672-1 and JIS C 1509-1 standards when the windscreen is installed.  Diffuse sound field correction:  Correction of frequency characteristics in order to comply with standards (ANSI S1.4) in diffuse sound field.  The meter can be set to start measuring a specified time (OFF, 1, 3, 5 or 10 s)
Switt RMS of Samp Calibr Correct	r graph detering detering ration	n display range max of bar graph display totion circuit cycle	Max. 110 dB (20 to 130 dB)  Set the upper/ lower limit in 10 dB increments.  Digital processing method  20.8 µs (Lp, Leq, Le, Lmax, Lmin, Lpeak : sampling frequency: 48 kHz)  100 ms (L <sub>N</sub> )  100 ms (L <sub>N</sub> )  Electrical calibration performed according to IEC and JIS standards, using internally generated signals: acoustic calibration performed with the NC-74.  Windscreen correction:  Compliant with IEC 61672-1 and JIS C 1509-1 standards when the windscreen is installed.  Diffuse sound field correction:  Correction of frequency characteristics in order to comply with standards (ANSI S1.4) in diffuse sound field.  The meter can be set to start measuring a specified time (OFF, 1, 3, 5 or 10 s) after the start button has been pressed or when a user-set trigger is exceeded.  When the PAUSE key is pressed to pause measurement, the preceding
Switt RMS 6 Samp Calibr Correct	r graph itching detection ration ration	n display range max of bar graph display totion circuit cycle	Max. 110 dB (20 to 130 dB)  Set the upper/ lower limit in 10 dB increments.  Digital processing method  20.8 μs (Lp, Leq, LE, Lmax, Lmin, Lpeak : sampling frequency: 48 kHz)  100 μs (Ln)  100 μs (Ln)  Insertal calibration performed according to IEC and JIS standards, using internally generated signals: acoustic calibration performed with the NC-74.  Windscreen correction:  Compliant with IEC 61672-1 and JIS C 1509-1 standards when the windscreen is installed.  Diffuse sound field correction:  Correction of frequency characteristics in order to comply with standards (ANSI S1.4) in diffuse sound field.  The meter can be set to start measuring a specified time (OFF, 1, 3, 5 or 10 s) after the start button has been pressed or when a user-set trigger is exceeded.  When the PAUSE key is pressed to pause measurement, the preceding (user selectable) 0, 1, 3 or 5 s data are excluded from processing.
Switt RMS 6 Samp Calibr Correct	r graph itching detection ration ration	n display range max of bar graph display totion circuit cycle	Max. 110 dB (20 to 130 dB)  Set the upper/ lower limit in 10 dB increments.  Digital processing method  20.8 μS (Lρ, Leq, LE, Lmax, Lmin, Lpeak : sampling frequency: 48 kHz)  100 ms (L <sub>N</sub> )  Electrical calibration performed according to IEC and JIS standards, using internally generated signals: acoustic calibration performed with the NC-74.  Windscreen correction:  Compliant with IEC 61672-1 and JIS C 1509-1 standards when the windscreen is installed.  Diffuse sound field correction:  Correction of frequency characteristics in order to comply with standards (ANSI S1.4) in diffuse sound field.  The meter can be set to start measuring a specified time (OFF, 1, 3, 5 or 10 s) after the start button has been pressed or when a user-set trigger is exceeded.  When the PAUSE key is pressed to pause measurement, the preceding (user selectable) 0, 1, 3 or 5 s data are excluded from processing.  Backlit semitransparent color TFT LCD display WQVGA (400 x 240 dots)
Switt RMS 6 Samp Calibr Correct	r graph itching detection ration ration	n display range max of bar graph display totion circuit cycle	Max. 110 dB (20 to 130 dB)  Set the upper/ lower limit in 10 dB increments.  Digital processing method  20.8 µS (Lp, Leq, LE, Lmax, Lmin, Lpeak : sampling frequency: 48 kHz)  100 ms (L <sub>N</sub> )  Electrical calibration performed according to IEC and JIS standards, using internally generated signals: acoustic calibration performed with the NC-74.  Windscreen correction:  Compliant with IEC 61672-1 and JIS C 1509-1 standards when the windscreen is installed.  Diffuse sound field correction:  Correction of frequency characteristics in order to comply with standards (ANSI S1.4) in diffuse sound field.  The meter can be set to start measuring a specified time (OFF, 1, 3, 5 or 10 s) after the start button has been pressed or when a user-set trigger is exceeded.  When the PAUSE key is pressed to pause measurement, the preceding (user selectable) 0, 1, 3 or 5 s data are excluded from processing.  Backlit semitransparent color TFT LCD display WQVGA (400 x 240 dots)  *LCD with touch panel (Capacitive Touch Panel)
Swith RMS (Samp Calibr Correct Delay Display	r graph teching deterior ration ration r time erase	n display range max of bar graph display ction circuit cycle functions	Max. 110 dB (20 to 130 dB)  Set the upper/ lower limit in 10 dB increments.  Digital processing method  20.8 µs (Lp, Leq, LE, Lmax, Lmin, Lpeak : sampling frequency: 48 kHz)  100 ms (Ln)  Electrical calibration performed according to IEC and JIS standards, using internally generated signals: acoustic calibration performed with the NC-74.  Windscreen correction:  Compliant with IEC 61672-1 and JIS C 1509-1 standards when the windscreen is installed.  Diffuse sound field correction:  Correction of frequency characteristics in order to comply with standards (ANSI S1.4) in diffuse sound field.  The meter can be set to start measuring a specified time (OFF, 1, 3, 5 or 10 s) after the start button has been pressed or when a user-set trigger is exceeded.  When the PAUSE key is pressed to pause measurement, the preceding (user selectable) 0, 1, 3 or 5 s data are excluded from processing.  Backlit semitransparent color TFT LCD display WQVGA (400 x 240 dots)  *LCD with touch panel (Capacitive Touch Panel)  Numerical display update frequency: 1 s Bar graph update frequency: 100 ms
Swith RMS (Samp Calibr Correct Delay Displa	r graph r graph detection ratior ratior rtime erase	n display range max of bar graph display totion circuit coycle  functions  e functions	Max. 110 dB (20 to 130 dB)  Set the upper/ lower limit in 10 dB increments.  Digital processing method  20.8 µs (Lp, Leq, LE, Lmax, Lmin, Lpeak : sampling frequency: 48 kHz)  100 ms (LN)  Electrical calibration performed according to IEC and JIS standards, using internally generated signals: acoustic calibration performed with the NC-74.  Windscreen correction:  Compliant with IEC 61672-1 and JIS C 1509-1 standards when the windscreen is installed.  Diffuse sound field correction:  Correction of frequency characteristics in order to comply with standards (ANSI S1.4) in diffuse sound field.  The meter can be set to start measuring a specified time (OFF, 1, 3, 5 or 10 s) after the start button has been pressed or when a user-set trigger is exceeded.  When the PAUSE key is pressed to pause measurement, the preceding (user selectable) 0, 1, 3 or 5 s data are excluded from processing.  Backlit semitransparent color TFT LCD display WQVGA (400 x 240 dots)  *LCD with touch panel (Capacitive Touch Panel)  Numerical display update frequency: 1 s Bar graph update frequency: 100 ms
Switt RMS (  RMS (	r graph r graph detection ratior ratior rtime erase	n display range max of bar graph display ction circuit cycle functions	Max. 110 dB (20 to 130 dB)  Set the upper/ lower limit in 10 dB increments.  Digital processing method  20.8 µs (L.p., Leq., L.E., Lmax, Lmin, Lpeak: sampling frequency: 48 kHz)  100 ms (L.N)  100 ms (L.N)  Electrical calibration performed according to IEC and JIS standards, using internally generated signals: acoustic calibration performed with the NC-74.  Windscreen correction:  Compliant with IEC 61672-1 and JIS C 1509-1 standards when the windscreen is installed.  Diffuse sound field correction:  Correction of frequency characteristics in order to comply with standards (ANSI S1.4) in diffuse sound field.  The meter can be set to start measuring a specified time (OFF, 1, 3, 5 or 10 s) after the start button has been pressed or when a user-set trigger is exceeded.  When the PAUSE key is pressed to pause measurement, the preceding (user selectable) 0, 1, 3 or 5 s data are excluded from processing.  Backlit semitransparent color TFT LCD display WQVGA (400 x 240 dots)  *LCD with touch panel (Capacitive Touch Panel)  Numerical display update frequency: 1 s Bar graph update frequency: 100 ms  Data for measurement results are stored manually in single address increments.
Swith RMS (Samp Calibr Correct Delay Displa	r graph detering detering ratior ratior ratior ration	n display range max of bar graph display cition circuit cycle  functions  functions	Max. 110 dB (20 to 130 dB)  Set the upper/ lower limit in 10 dB increments.  Digital processing method  20.8 μS (Lρ, Leq, LE, Lmax, Lmin, Lpeak : sampling frequency: 48 kHz)  100 ms (L <sub>N</sub> )  Electrical calibration performed according to IEC and JIS standards, using internally generated signals: acoustic calibration performed with the NC-74.  Windscreen correction:  Compliant with IEC 61672-1 and JIS C 1509-1 standards when the windscreen is installed.  Diffuse sound field correction:  Correction of frequency characteristics in order to comply with standards (ANSI S1.4) in diffuse sound field.  The meter can be set to start measuring a specified time (OFF, 1, 3, 5 or 10 s) after the start button has been pressed or when a user-set trigger is exceeded.  When the PAUSE key is pressed to pause measurement, the preceding (user selectable) 0, 1, 3 or 5 s data are excluded from processing.  Backlit semitransparent color TFT LCD display WQVGA (400 x 240 dots) *LCD with touch panel (Capacitive Touch Panel)  Numerical display update frequency: 1 s Bar graph update frequency: 100 ms  Data for measurement results are stored manually in single address increments.  Internal memory: max. 1000 sets  SD Card: depends on the capacity of the SD Card*1
Swith RMS (Samp Calibr Correct Delay Displa	r graph r graph detection ratior ratior rtime erase	n display range max of bar graph display cition circuit cycle  functions  functions	Max. 110 dB (20 to 130 dB)  Set the upper/ lower limit in 10 dB increments.  Digital processing method  20.8 µs (Lp, Leq, LE, Lmax, Lmin, Lpeak : sampling frequency: 48 kHz)  100 ms (Ln)  101 ms (Ln)  101 ms (Ln)  102 ms (Ln)  103 ms (Ln)  103 ms (Ln)  104 ms (Ln)  105 ms (Ln)  105 ms (Ln)  106 ms (Ln)  107 ms (Ln)  108 ms (Ln)  109 ms (Ln)  109 ms (Ln)  109 ms (Ln)  100 ms (Ln)
Swith RMS (Samp Calibr Correct Delay Displa	r graph detering detering ratior ratior ratior ration	n display range max of bar graph display cition circuit cycle  functions  functions	Max. 110 dB (20 to 130 dB)  Set the upper/ lower limit in 10 dB increments.  Digital processing method  20.8 μS (Lρ, Leq, LE, Lmax, Lmin, Lpeak : sampling frequency: 48 kHz)  100 ms (L <sub>N</sub> )  Electrical calibration performed according to IEC and JIS standards, using internally generated signals: acoustic calibration performed with the NC-74.  Windscreen correction:  Compliant with IEC 61672-1 and JIS C 1509-1 standards when the windscreen is installed.  Diffuse sound field correction:  Correction of frequency characteristics in order to comply with standards (ANSI S1.4) in diffuse sound field.  The meter can be set to start measuring a specified time (OFF, 1, 3, 5 or 10 s) after the start button has been pressed or when a user-set trigger is exceeded.  When the PAUSE key is pressed to pause measurement, the preceding (user selectable) 0, 1, 3 or 5 s data are excluded from processing.  Backlit semitransparent color TFT LCD display WQVGA (400 x 240 dots) *LCD with touch panel (Capacitive Touch Panel)  Numerical display update frequency: 1 s Bar graph update frequency: 100 ms  Data for measurement results are stored manually in single address increments.  Internal memory: max. 1000 sets  SD Card: depends on the capacity of the SD Card*1
Swith RMS (Samp Calibr Correct Delay Displa	r graph r graph r graph r graph r graph r graph determ ration r time erase  Maa  Aut	n display range max of bar graph display cition circuit cycle  functions  functions	Max. 110 dB (20 to 130 dB)  Set the upper/ lower limit in 10 dB increments.  Digital processing method  20.8 μS (Lρ, Leq, LE, Lmax, Lmin, Lpeak : sampling frequency: 48 kHz)  100 ms (L <sub>N</sub> )  Electrical calibration performed according to IEC and JIS standards, using internally generated signals: acoustic calibration performed with the NC-74.  Windscreen correction:  Compliant with IEC 61672-1 and JIS C 1509-1 standards when the windscreen is installed.  Diffuse sound field correction:  Correction of frequency characteristics in order to comply with standards (ANSI S1.4) in diffuse sound field.  The meter can be set to start measuring a specified time (OFF, 1, 3, 5 or 10 s) after the start button has been pressed or when a user-set trigger is exceeded.  When the PAUSE key is pressed to pause measurement, the preceding (user selectable) 0, 1, 3 or 5 s data are excluded from processing.  Backlit semitransparent color TFT LCD display WQVGA (400 x 240 dots) *LCD with touch panel (Capacitive Touch Panel)  Numerical display update frequency: 1 s Bar graph update frequency: 100 ms  Data for measurement results are stored manually in single address increments. Internal memory: max. 1000 sets  SD Card: depends on the capacity of the SD Card*1  Instantaneous values (Lρ mode) and processed values (Leq mode) are
Swith RMS (Samp Calibr Correct Delay Displa	r graph r graph r graph r graph r graph r determined de	n display range max of bar graph display totion circuit cycle  functions  functions  e function	Max. 110 dB (20 to 130 dB)  Set the upper/ lower limit in 10 dB increments.  Digital processing method  20.8 µs (Lp, Leq, LE, Lmax, Lmin, Lpeak : sampling frequency: 48 kHz)  100 ms (L <sub>N</sub> )  Electrical calibration performed according to IEC and JIS standards, using internally generated signals: acoustic calibration performed with the NC-74.  Windscreen correction:  Compliant with IEC 61672-1 and JIS C 1509-1 standards when the windscreen is installed.  Diffuse sound field correction:  Correction of frequency characteristics in order to comply with standards (ANSI S1.4) in diffuse sound field.  The meter can be set to start measuring a specified time (OFF, 1, 3, 5 or 10 s) after the start button has been pressed or when a user-set trigger is exceeded.  When the PAUSE key is pressed to pause measurement, the preceding (user selectable) 0, 1, 3 or 5 s data are excluded from processing.  Backlit semitransparent color TFT LCD display WOVGA (400 x 240 dots)  *LCD with touch panel (Capacitive Touch Panel)  Numerical display update frequency: 1 s Bar graph update frequency: 100 ms  Data for measurement results are stored manually in single address increments.  Internal memory: max. 1000 sets  SD Card: depends on the capacity of the SD Card*1  Instantaneous values (Lp mode) and processed values (Leq mode) are stored continuously and automatically at preset intervals.
Switt RMS of Samp Calibr Correct	rgraph graph	n display range max of bar graph display cition circuit cycle  functions  functions  definitions  are function	Max. 110 dB (20 to 130 dB)  Set the upper/ lower limit in 10 dB increments.  Digital processing method  20.8 µs (Lp, Leq, LE, Lmax, Lmin, Lpeak : sampling frequency: 48 kHz)  100 ms (LN)  Electrical calibration performed according to IEC and JIS standards, using internally generated signals: acoustic calibration performed with the NC-74.  Windscreen correction:  Compliant with IEC 61672-1 and JIS C 1509-1 standards when the windscreen is installed.  Diffuse sound field correction:  Correction of frequency characteristics in order to comply with standards (ANSI S1.4) in diffuse sound field.  The meter can be set to start measuring a specified time (OFF, 1, 3, 5 or 10 s) after the start button has been pressed or when a user-set trigger is exceeded.  When the PAUSE key is pressed to pause measurement, the preceding (user selectable) 0, 1, 3 or 5 s data are excluded from processing.  Sacklit semitransparent color TFT LCD display WQVGA (400 x 240 dots)  *LCD with touch panel (Capacitive Touch Panel)  Numerical display update frequency: 1 s Bar graph update frequency: 100 ms  Data for measurement results are stored manually in single address increments. Internal memory: max. 1000 sets  SD Card: dependent and user selected time (up to 24 hours)  100 ms, 200 min, 1 s, Leq 1s and user selected time (up to 24 hours)

Da	ta recall		Allows viewing of stored data	
Set	tup men	nory	Up to five setup configurations can be saved in internal memory, for later recall	
			Start up via file settings previously stored on SD card possible	
Wa	veform	recording*2		
File format		nat	Uncompressed waveform WAVE file	
Sampling frequency		g frequency	Select 48 kHz, 24 kHz or 12 kHz	
Data length		ngth	Select 24 bit or 16 bit	
Outputs DC output		output	Output DC signals using a frequency weighting characteristic selected by processing	
	C	Output voltage	2.5 V, 25 mV / dB at bar graph display full scale	
	AC	output	Output AC signal using frequency weighting selected by processing or by A	
			C, Z, G weighting	
	C	Output voltage	1 V (rms values) at bar graph display full scale	
	Cor	nparator	Turns on when the open-collector output exceeds the set value	
	out	out	(max. applied voltage 24 V, max. current 60 mA, allowable dissipation 300 mW).	
US	В		Allows USB to be connected to a computer and recognized as a removable disl	
			Allows USB to be controlled via communication commands	
RS	-232C c	ommunication	Allows for RS-232C communication via use of a dedicated cable	
Da	ta contin	uous output		
	Type of	Instantaneous value	Lp	
	data	Processed value	Leq, Lmax, Lmin, Lpeak	
Ī	Output i	nterval	100 ms	
Pri	nt out		Printing of measurement results on dedicated printer DPU-414	
Po	wer requ	uirements	Four IEC R6 (size AA) batteries (alkaline or rechargeable batteries) or external power supply	
	Battery	life (23 °C)	Alkaline battery LR6 (AA): 16 h Ni-MH secondary battery: 16 h	
L			At the maximum *Depends on the setting	
	AC ada		NC-98C	
L	External	power voltage	5 to 7 V (rated voltage: 6 V)	
	Current	consumption	Approximately 120 mA (normal operation, rated voltage)	
Am	bient	Temperature	-10 to +50 °C	
cor	nditions	Humidity	10 to 90 % RH (non-condensing)	
Du	stproof /	water-resistant	IP code: IP54 (except for microphone)	
performance*3 See precautions regarding waterproofing		See precautions regarding waterproofing		
Din	nension	s, weight	Approx. 255 (H) x 76 (W) x 33 mm(D), approx. 400 g (with batteries)	
Su	pplied a	ccessories	Storage case x 1, Windscreen WS-10 x 1, Windscreen fall prevention rubber x 1,	
			Hand strap x 1, LR6 (AA) alkaline batteries x 4, SD card 512 MB×1	

#### Ontions

Product name	Product number	
Waveform recording program (Inst.on 2 GB SD card)	NX-42WR	
Octave, 1/3 octave real-time analysis program (Inst.on 512 MB SD card)	NX-62RT	
Reverberation time measurement program (Inst.on 512 MB SD card)	NX-42RV	
FFT analysis program (Inst.on 512 MB SD card)	NX-42FT	
Data management software for environmental measurement	AS-60	
Data management software for environmental measurement (Includes the octave and 1/3 octave data management software)	AS-60RT	
Data management software for environmental measurement (Includes the vibration level data management software)	AS-60VM	
Waveform analysis software	AS-70	
SD Card 512 MB	MC-51SD1	
SD Card 2 GB	MC-20SD2	
AC adapter (100 V to 240 V)	NC-98C	
Battery pack	BP-21A	
Microphone extension cables	EC-04 (from 2 m)	
BNC-Pin output code	CC-24	
Comparator output cable	CC-42C	
Printer	DPU-414	
Printer cable	CC-42P	
RS 232C serial I/O cable	CC-42R	
USB cable	Generic USB cable can be used	
Sound calibrator	NC-74	
All-weather windscreen	WS-15	
Windscreen mounting adapter	WS-15006	
Rain-protection windscreen	WS-16	
Sound level meter tripod	ST-80	
All-weather windscreen tripod	ST-81	

- \*1 Use Rion fully guaranteed products. \*2 NX-42WR required (sold separately). st3 Protection against harmful dust and water splashing from any direction.

Precautions regarding waterproofing
Before use, verify that the rubber bottom cover and the battery compartment lid are firmly closed. To maintain the water and dust proof rating, internal packing replacement is required every two years (at cost).

ISO 14001 RION CO., LTD. ISO 9001 RION CO., LTD.

- \* Windows is a trademark of Microsoft Corporation.
- \* Specifications subject to change without notice.

Distributed by:



3-20-41, Higashimotomachi, Kokubunji, Tokyo 185-8533, Japan Tel: +81-42-359-7888 Fax: +81-42-359-7442